

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

	ed Product: VUL385 COPPER	R STANDARD 5.0 PPM SOLU	TION	
Test		Target/UOM	Range	Result
	D TO FORMULATION ON	YES		YES
FILE		YES		
	Sample ID			
	Tested 04/25/24			
	Tester MMAT			
	Recorded 04/25/24			
PPM CU		5.00	4.95 - 5.05	5.00
		PPM		
	Sample ID			
	Tested 04/25/24			
	Tester MMAT			
	Recorded 04/25/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
	Sample ID			
	Tested 04/25/24			
	Tester MMAT			
	Recorded 04/25/24			
TRACEABLE TO N.I.S.T. STANDARD		YES		YES
		YES		
	Sample ID			
	Tested 04/25/24			
	Tester MMAT			
	Recorded 04/25/24			
	ENTS USED DURING	INSTRUMENT		P69,GC,VF4-N30
PREPARA				
	Sample ID			
	Tested 04/25/24			
	Tester MMAT			
	Recorded 04/25/24			
Actual Der	nsity	1.00344	1.00314 - 1.00374	1.00329
		GM/ML		
	Sample ID			
	Tested 04/25/24			
	Tester MMAT			
	Recorded 04/25/24			40/00/04
EXPIRATIO	JNUAIE			10/30/24
	O annula ID	MM/DD/YYYY		
	Sample ID			
	Tested 04/25/24			
	Tester MMAT			
	Recorded 04/25/24			
Lo	t# 4042483			
Ma	de 04/25/24			



Page: 2 Date: 04/25/24 at 5:49 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager