Page: 1

Date: 04/26/24 at 1:18 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	Sodium Standard Ref1 5.0 ppm for EZ3015-Sodium Analyzer				
Test			Target/UOM	Range	Result
	ED TO FORMU	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	04/26/24			
	Tester	JTRA			
	Recorded	04/26/24			
APPEARANCE AND COLOR		LOR	CLEAR WATER WHITE		PASS
			APPEARANCE		
	Sample ID				
	Tested	04/26/24			
	Tester	JTRA			
	Recorded	04/26/24			
NSTRUM	ENTS USED D	URING	INSTRUMENT		P94,VF2-B3
PREPARA	ATION				
	Sample ID				
	Tested	04/26/24			
	Tester	JTRA			
	Recorded	04/26/24			
EXPIRATION DATE			MM/DD/YY		07/30/24
			MM/DD/YYYY		
	Sample ID				
	Tested	04/26/24			
	Tester	JTRA			
	Recorded	04/26/24			
РРМ	. 1000. 000	0 1/20/21	5.00	4.95 - 5.05	5.00
			PPM		
	Sample ID				
	Tested	04/26/24			
	Tester	04/26/24 JTRA			
	Recorded	04/26/24			
	ot # 40426				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland In