



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: PPG036F COLOR REAGENT (CHLORIDES)			
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 04/24/24			
Tester MMAT			
Recorded 04/24/24			
Final Product Filtered	YES		YES
Sample ID	YES		
Tested 04/24/24			
Tester MMAT			
Recorded 04/24/24			
Renex 30, 30% added to product	YES		YES
Sample ID	YES		
Tested 04/24/24			
Tester MMAT			
Recorded 04/24/24			
ACTUAL GRAMS OF MERCURIC THIOCYANATE	RECORD ACTUAL GMS M2275 GM		0.5200
Sample ID			
Tested 04/24/24			
Tester MMAT			
Recorded 04/24/24			
RECORD ACTUAL MLS OF METHANOL USED	RECORD ACTUAL MLS OF METHANOL MLS		200
Sample ID			
Tested 04/24/24			
Tester MMAT			
Recorded 04/24/24			
TOTAL MLS OF NITRIC ACID ADDED	RECORD ACTUAL MLS NITRIC ACID MLS		63
Sample ID			
Tested 04/24/24			
Tester MMAT			
Recorded 04/24/24			
RECORD ACTUAL GRAMS OF FERRIC NITRATE	RECORD ACT GMS FERRIC NITRATE GMS F2530		40.00
Sample ID			
Tested 04/24/24			
Tester MMAT			
Recorded 04/24/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

FINAL VOLUME IN LITERS	RECORD FINAL VOLUME IN	1
	LITERS	
	LITER	
Sample ID		
Tested 04/24/24		
Tester MMAT		
Recorded 04/24/24		
APPEARANCE AND COLOR	BROWN LIQUID	PASS
	APPEARANCE	
Sample ID		
Tested 04/24/24		
Tester MMAT		
Recorded 04/24/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M115,GC,VF1-Z44,M100
Sample ID		
Tested 04/24/24		
Tester MMAT		
Recorded 04/24/24		
VERIFICATION	PASS	PASS
	PASS	
Sample ID		
Tested 04/24/24		
Tester MMAT		
Recorded 04/24/24		
EXPIRATION DATE	MM/DD/YYYY	10/30/2024
	MM/DD/YYYY	
Sample ID		
Tested 04/24/24		
Tester MMAT		
Recorded 04/24/24		
Lot #	4042401	
Made	04/24/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager