

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

	Product:	SODIUM	CARBONATE 1.0 NORMAL (0	.5 Molar)	
Test			Target/UOM	Range	Result
Actual Norm	mality		1.0000	0.9995 - 1.0005	0.9997
			EQ/L		
	Sample ID				
		/22/24			
		NOR			
		/22/24			
STANDARD	IZED USING QC	SP36	YES		YES
			PASS		
	Sample ID				
		/22/24			
		NOR			
		/22/24			
N.I.S.T. Traceable to SRM 84		4	Yes		YES
			YES		
	Sample ID				
		/22/24			
		NOR			
		/22/24			
	TO FORMULAT	TON ON	YES		YES
FILE			YES		
	Sample ID				
	Tested 04	/22/24			
	Tester EN	NOR			
	Recorded 04/	/22/24			
Actual Density			1.05097	1.05048 - 1.05246	1.05071
			GM/ML		
	Sample ID				
	Tested 04	/22/24			
	Tester EN	NOR			
		/22/24			
APPEARANCE AND COLOR		2	CLEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested 04	/22/24			
		NOR			
	Recorded 04	/22/24			
INSTRUMEN	NTS USED DURI	ING	INSTRUMENT		M112,T-190,BID-08446,PID-69
PREPARATI	ION				
	Sample ID				
	Tested 04	/22/24			
	Tester EN	NOR			
		/22/24			
EXPIRATIO	N DATE		MM/DD/YY		04/30/25
			MM/DD/YYYY		
	Sample ID				
	-	/22/24			
	Tester EN	NOR			
	Recorded 04/	/22/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4041956
Made	04/22/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

thand St

Quality Manager