

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	ND334F			
		JTION pH 9.6 (N.I.S.T.		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	04/25/24			
Tester	JRIO			
Recorded	04/25/24			
APPEARANCE, COLOR:	CLEA	R COLORLESS LIQUID		PASS
		COLOR		
Sample ID				
Tested	04/25/24			
Tester	JRIO			
Recorded	04/25/24			
NSTRUMENTS USED D	JRING	INSTRUMENT		M103,EF2-COCO
PREPARATION				
Sample ID				
Tested	04/25/24			
Tester	JRIO			
Recorded	04/25/24			
PH OF PRODUCT/SOLUTION		9.600	9.580 - 9.620	9.601
		PH		
Sample ID				
Tested	04/25/24			
Tester	JRIO			
Recorded	04/25/24			
EXPIRATION DATE		MM/DD/YY		04/30/25
		MM/DD/YYYY		
Sample ID				
Tested	04/25/24			
Tester	JRIO			
Recorded	04/25/24			
Lot # 40425	51			
Tester Recorded Lot # 40425	JRIO 04/25/24 61			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA