



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7320			
POTASSIUM IODATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.1000	0.0995 - 0.1005	0.0996
	EQ/L		
Sample ID			
Tested	04/25/24		
Tester	JENC		
Recorded	04/25/24		
QCSP33:POTASSIUM IODATE-IODIDE TITRANT	YES		YES
	QCSP33		
Sample ID			
Tested	04/25/24		
Tester	JENC		
Recorded	04/25/24		
Traceable to N.I.S.T. SRM 185	Yes		YES
	YES		
Sample ID			
Tested	04/25/24		
Tester	JENC		
Recorded	04/25/24		
Actual Density	1.00171	1.00151 - 1.00191	1.00175
	GM/ML		
Sample ID			
Tested	04/25/24		
Tester	JENC		
Recorded	04/25/24		
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested	04/25/24		
Tester	JENC		
Recorded	04/25/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	04/25/24		
Tester	JENC		
Recorded	04/25/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,T44
Sample ID			
Tested	04/25/24		
Tester	JENC		
Recorded	04/25/24		
EXPIRATION DATE	MM/DD/YYYY		04/30/2025
	MM/DD/YYYY		
Sample ID			
Tested	04/25/24		
Tester	JENC		
Recorded	04/25/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 04/25/24 at 2:04 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	4042447
Made	04/25/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager