

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Package	ed Product	6494			
		NITROG	EN STD 100 PPM ($1ML = 0$.	1MG N) SOLUTION	
Test			Target/UOM	Range	Result
PREPARE	D TO FORMU	ILATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	04/23/24			
	Tester	MMAT			
	Recorded	04/23/24			
CERTIFIED PPM NITROGEN:		GEN:	100.0	96.0 - 104.0	100.4
			PPM		
	Sample ID				
	Tested	04/23/24			
	Tester	MMAT			
	Recorded	04/23/24			
APPEARANCE AND COLOR			CLEAR WATER WHITE		PASS
		-	LIQUID		
			APPEARANCE		
	Sample ID				
	Tested	04/23/24			
	Tester	MMAT			
	Recorded	04/23/24			
	ENTS USED E		INSTRUMENT		M108,VF4-N29
PREPARA			ING INCOMENT		10100,01 4-1023
FREFARA					
	Sample ID Tested	04/00/04			
		04/23/24			
	Tester	MMAT			
	Recorded	04/23/24			
EXPIRATI	ON DATE		MM/DD/YYYY		M108,GP,VF4-N29
1			MM/DD/YYYY		
	Sample ID				
	Tested	04/23/24			
	Tester	MMAT			
	Recorded	04/23/24			
Lo	ot# 40422	246			
Ма	de 04/23	3/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1