



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: VEN004F			
Mixed ICP Standard 10.0 PPM in 5% nitric			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 04/29/24			
Tester MMAT			
Recorded 04/29/24			
CALCULATED PPM CALCIUM	10.00 PPM	9.90 - 10.10	10.00
Sample ID			
Tested 04/29/24			
Tester MMAT			
Recorded 04/29/24			
CALCULATED PPM COPPER (CU)	10.00 PPM	9.90 - 10.10	10.00
Sample ID			
Tested 04/29/24			
Tester MMAT			
Recorded 04/29/24			
PPM IRON (FE)	10.00 PPM	9.90 - 10.10	10.00
Sample ID			
Tested 04/29/24			
Tester MMAT			
Recorded 04/29/24			
CALCULATED PPM POTASSIUM	10.00 MG/L	9.90 - 10.10	10.00
Sample ID			
Tested 04/29/24			
Tester MMAT			
Recorded 04/29/24			
CALCULATED PPM MAGNESIUM (MG)	10.00 PPM	9.90 - 10.10	10.00
Sample ID			
Tested 04/29/24			
Tester MMAT			
Recorded 04/29/24			
CALCULATED PPM MANGANESE	10.00 PPM (UG/ML)	9.90 - 10.10	10.00
Sample ID			
Tested 04/29/24			
Tester MMAT			
Recorded 04/29/24			
PPM SODIUM	10.00 PPM	9.90 - 10.10	10.00
Sample ID			
Tested 04/29/24			
Tester MMAT			
Recorded 04/29/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

PPM OF LEAD	10.00	9.90 - 10.10	10.00
	PPM		
Sample ID			
Tested	04/29/24		
Tester	MMAT		
Recorded	04/29/24		
CALCULATED MG/L SILICA (SI)	10.00	9.90 - 10.10	10.00
	MG/L		
Sample ID			
Tested	04/29/24		
Tester	MMAT		
Recorded	04/29/24		
CALCULATED PPM ZINC (ZN)	10.00	9.90 - 10.10	10.00
	PPM		
Sample ID			
Tested	04/29/24		
Tester	MMAT		
Recorded	04/29/24		
APPEARANCE AND COLOR	PASS		PASS
	APPEARANCE		
Sample ID			
Tested	04/29/24		
Tester	MMAT		
Recorded	04/29/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P89,VF1-Z45,GC
Sample ID			
Tested	04/29/24		
Tester	MMAT		
Recorded	04/29/24		
EXPIRATION DATE	MM/DD/YYYY		04/30/2025
	MM/DD/YYYY		
Sample ID			
Tested	04/29/24		
Tester	MMAT		
Recorded	04/29/24		
Lot #	4041844		
Made	04/29/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 04/29/24 at 11:01 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager