



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: SPX327F			
	SODIUM CARBONATE 1% W/V; SODIUM BICARBONATE 1% W/V		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 05/03/24			
Tester JTRA			
Recorded 05/03/24			
ACT GMS SODIUM CARBONATE	10.0000 GMS	9.9975 - 10.0025	9.9999
Sample ID			
Tested 05/03/24			
Tester JTRA			
Recorded 05/03/24			
ACTUAL GRAMS SODIUM BICARBONATE	10.0000 GMS	9.9600 - 10.0200	9.9900
Sample ID			
Tested 05/03/24			
Tester JTRA			
Recorded 05/03/24			
ASSAY OF SODIUM CARBONATE	100.00 W/W%	99.95 - 100.05	100.01
Sample ID			
Tested 05/03/24			
Tester JTRA			
Recorded 05/03/24			
ASSAY SODIUM BICARBONATE	100.0 WT%	98.5 - 101.0	100.1
Sample ID			
Tested 05/03/24			
Tester JTRA			
Recorded 05/03/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 05/03/24			
Tester JTRA			
Recorded 05/03/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M108,VF1-Z20
Sample ID			
Tested 05/03/24			
Tester JTRA			
Recorded 05/03/24			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		08/30/2024
Sample ID			
Tested 05/03/24			
Tester JTRA			
Recorded 05/03/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 05/03/24 at 2:25 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	4050355
Made	05/03/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager