



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8365			
SILVER NITRATE 0.001 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 04/29/24			
Tester JRIO			
Recorded 04/29/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID	YES		
Tested 04/29/24			
Tester JRIO			
Recorded 04/29/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		04/29/2024
Sample ID	MM/DD/YYYY		
Tested 04/29/24			
Tester JRIO			
Recorded 04/29/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	20.3
Sample ID	DEG C		
Tested 04/29/24			
Tester JRIO			
Recorded 04/29/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		04/29/2024
Sample ID	MM/DD/YYYY		
Tested 04/29/24			
Tester ENOR			
Recorded 04/29/24			
<b>AMBIENT TEST TEMPERATURE DEG C</b>	20.0	15.0 - 30.0	20.0
Sample ID	DEG C		
Tested 04/29/24			
Tester ENOR			
Recorded 04/29/24			
<b>QCSP54:STANDARDIZATION OF LOW NORMAL SN</b>	YES		YES
Sample ID	QCSP54		
Tested 04/29/24			
Tester ENOR			
Recorded 04/29/24			
<b>N.I.S.T. Traceable to SRM 919 NaCl</b>	YES		YES
Sample ID	YES		
Tested 04/29/24			
Tester ENOR			
Recorded 04/29/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE	PASS
	LIQUID APPEARANCE	
Sample ID		
Tested 04/29/24		
Tester ENOR		
Recorded 04/29/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>INSTRUMENT</b>	GC,VF4-N6,BID-AT,PID-160
Sample ID		
Tested 04/29/24		
Tester ENOR		
Recorded 04/29/24		
<b>NORMALITY TO 5 DECIMAL PLACES</b>	0.00100	0.00080 - 0.00120
	EQUIV/LITER	0.00097
Sample ID		
Tested 04/29/24		
Tester ENOR		
Recorded 04/29/24		
<b>EXPIRATION DATE</b>	MM/DD/YYYY	04/30/2025
	MM/DD/YYYY	
Sample ID		
Tested 04/29/24		
Tester ENOR		
Recorded 04/29/24		
<b>Lot #</b>	4042528	
<b>Made</b>	04/29/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**