



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> EP035F			
	SODIUM ISOPROPOXIDE 0.1 NORMAL (MAKE FRESH)		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 05/03/24			
Tester ENOR			
Recorded 05/03/24			
<b>Actual Normality</b>	0.1000	0.0995 - 0.1005	0.0999
	EQ/L		
Sample ID			
Tested 05/03/24			
Tester ENOR			
Recorded 05/03/24			
<b>QCSP18:STANDARDIZATION</b>	YES		YES
<b>SODIUM ISOPROPROX</b>	QCSP18		
Sample ID			
Tested 05/03/24			
Tester ENOR			
Recorded 05/03/24			
<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES		YES
	YES		
Sample ID			
Tested 05/03/24			
Tester ENOR			
Recorded 05/03/24			
<b>Actual Density</b>	0.79146	0.79085 - 0.79206	0.79164
	GM/ML		
Sample ID			
Tested 05/03/24			
Tester ENOR			
Recorded 05/03/24			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 05/03/24			
Tester ENOR			
Recorded 05/03/24			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M112,VF12-9,VF3-J3,GC,M55
Sample ID			
Tested 05/03/24			
Tester ENOR			
Recorded 05/03/24			
<b>SHELF LIFE OF PRODUCT</b>	9 MONTHS		9 MONTHS
	TIME		
Sample ID			
Tested 05/03/24			
Tester ENOR			
Recorded 05/03/24			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>EXPIRATION DATE</b>	MM/DD/YYYY	02/28/2025
	MM/DD/YYYY	
Sample ID		
Tested	05/03/24	
Tester	ENOR	
Recorded	05/03/24	
<b>Lot #</b>	4042640	
<b>Made</b>	05/03/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**