

Page: 1

Date: 05/09/24 at 4:13 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	4548			
	INORGAN	IC CARBON STANDARD 10.0	ppm, Certified	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested 0	05/09/24			
Tester [OTRA			
Recorded 0	05/09/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
· ·	05/09/24			
	OTRA			
	05/09/24			
INSTRUMENTS USED DU		INSTRUMENT		P98,VF1-Z33
PREPARATION				,
Sample ID				
· ·	05/09/24			
	OTRA			
	05/09/24			
EXPIRATION DATE	33/03/24	MM/DD/YY		07/14/24
EXI IIIATION DATE		MM/DD/YYYY		07711721
Sample ID		WIWI, DD, TTTT		
·	05/09/24			
	05/09/24 DTRA			
	05/09/24			
SHELF LIFE OF PRODUC		SHELF LIFE		2 MONTHS
SHELF LIFE OF PRODUC	•	TIME		2 100111113
Commis ID		IIIVIL		
Sample ID Tested 0	25/00/04			
	05/09/24			
	OTRA			
	05/09/24	10.00	0.60 40.40	0.72
IC TO 2 DECIMAL PLACES	3	PPM	9.60 - 10.40	9.73
2		PPIVI		
Sample ID	= 10010:			
	05/09/24			
	OTRA			
	05/09/24			
Lot # 405094	5			
Made 05/09/	24			

Page: 2

Date: 05/09/24 at 4:13 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager