



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 1527			
BUFFER REFERENCE pH 12.0 (±0.02)			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 05/07/24			
Tester MMAT			
Recorded 05/07/24			
<b>PH OF PRODUCT/SOLUTION</b>	12.000	11.980 - 12.020	11.998
	PH		
Sample ID			
Tested 05/07/24			
Tester JRIO			
Recorded 05/07/24			
<b>QCSP 20 PH CALIBRATION METHODS</b>	YES		YES
	METHOD		
Sample ID			
Tested 05/07/24			
Tester JRIO			
Recorded 05/07/24			
<b>Actual Temp</b>	25.0	24.0 - 26.0	25.0
	DEG C		
Sample ID			
Tested 05/07/24			
Tester JRIO			
Recorded 05/07/24			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 05/07/24			
Tester JRIO			
Recorded 05/07/24			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M103,T21
Sample ID			
Tested 05/07/24			
Tester JRIO			
Recorded 05/07/24			
<b>PH NIST TRACEABLE TO SRM 187</b>	YES		YES
	YES		
Sample ID			
Tested 05/07/24			
Tester JRIO			
Recorded 05/07/24			
<b>PH NIST TRACEABLE TO SRM 2193</b>	YES		YES
	YES		
Sample ID			
Tested 05/07/24			
Tester JRIO			
Recorded 05/07/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>TRACEABLE TO NIST SRM 191I AND SRM 191II</b>	YES YES	YES
Sample ID		
Tested	05/07/24	
Tester	JRIO	
Recorded	05/07/24	
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	05/30/25
Sample ID		
Tested	05/07/24	
Tester	JRIO	
Recorded	05/07/24	
<b>SHELF LIFE OF PRODUCT</b>	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested	05/07/24	
Tester	JRIO	
Recorded	05/07/24	
<b>Lot #</b>	4042473	
<b>Made</b>	05/07/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



**Quality Manager**