

Page: 1

Date: 05/09/24 at 3:16 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 6991			
CONDUC	CTIVITY STANDARD 65,000	UMHOS	
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE			
Sample ID			
Tested 05/09/24			
Tester JRIO			
Recorded 05/09/24			
CONDUCTIVITY @25 C	65000	64900 - 65100	65000
	US/CM		
Sample ID			
Tested 05/09/24			
Tester JRIO			
Recorded 05/09/24			
QCSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 05/09/24			
Tester JRIO			
Recorded 05/09/24			
TRACEABLE TO N.I.S.T. SRM 3191	YES		YES
	YES		
Sample ID			
Tested 05/09/24			
Tester JRIO			
Recorded 05/09/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested 05/09/24			
Tester JRIO			
Recorded 05/09/24			
INSTRUMENTS USED DURING	INSTRUMENT		M92,T234
PREPARATION			·
Sample ID			
Tested 05/09/24			
Tester JRIO			
Recorded 05/09/24			
EXPIRATION DATE	MM/DD/YYYY		05/30/2025
	MM/DD/YYYY		
Sample ID			
Tested 05/09/24			
Tester JRIO			
Recorded 05/09/24			
Lot # 4050986			
Made 05/09/24			
Made 00/09/24			



Page: 2

Date: 05/09/24 at 3:16 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager