



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: SPX376F | | | |
|--|-------------------------------------|-----------|-----------------|
| TOC STANDARD 375PPM (KHP) | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | YES | | |
| Tested 05/07/24 | | | |
| Tester DTRA | | | |
| Recorded 05/07/24 | | | |
| PPM TOTAL ORGANIC CARBON | 375 | 369 - 381 | 372 |
| | MG/L (PPM) | | |
| Sample ID | | | |
| Tested 05/07/24 | | | |
| Tester DTRA | | | |
| Recorded 05/07/24 | | | |
| APPEARANCE AND COLOR | CLEAR WATER WHITE LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested 05/07/24 | | | |
| Tester DTRA | | | |
| Recorded 05/07/24 | | | |
| PRESERVED WITH HYDROCHLORIC ACID | YES | | YES |
| | YES | | |
| Sample ID | | | |
| Tested 05/07/24 | | | |
| Tester DTRA | | | |
| Recorded 05/07/24 | | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | M108,GC,VF1-Z16 |
| Sample ID | | | |
| Tested 05/07/24 | | | |
| Tester DTRA | | | |
| Recorded 05/07/24 | | | |
| EXPIRATION DATE | MM/DD/YY | | 11/30/24 |
| | MM/DD/YYYY | | |
| Sample ID | | | |
| Tested 05/07/24 | | | |
| Tester DTRA | | | |
| Recorded 05/07/24 | | | |
| Lot # | 4050732 | | |
| Made | 05/07/24 | | |



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536

Website: www.aquasolutions.org

Page: 2

Date: 05/07/24 at 2:49 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager