05/09/24

Made

Page: 1

Date: 05/09/24 at 2:26 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## Certificate of Analysis / QC Results

Customer PO #:

6490 Packaged Product: NITROGEN STANDARD 1000 PPM (1ML=1MG N), CERTIFIED Target/UOM Range Result Test PREPARED TO FORMULATION ON YES YES YES **FILE** Sample ID Tested 05/09/24 Tester **MMAT** 05/09/24 Recorded PPM NITROGEN NO DECIAML 1000 980 - 1020 1012 **PLACES** PPM Sample ID Tested 05/09/24 Tester **MMAT** Recorded 05/09/24 APPEARANCE AND COLOR **CLEAR WATER WHITE PASS** LIQUID **APPEARANCE** Sample ID Tested 05/09/24 Tester **MMAT** 05/09/24 Recorded M108, VF1-Z42 INSTRUMENTS USED DURING INSTRUMENT **PREPARATION** Sample ID Tested 05/09/24 Tester MMAT Recorded 05/09/24 **EXPIRATION DATE** MM/DD/YY 08/30/24 MM/DD/YYYY Sample ID Tested 05/09/24 Tester **MMAT** 05/09/24 Recorded 4050936 Lot#

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1.