

Recorded 05/08/24

Page: 1

Date: 05/08/24 at 1:32 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Customer PO # :				
Packaged Product:	6760			
•	PHOSPHA	TE BUFFER PH 7.2 A.P.H.	A. FOR BOD	
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES	Range	YES
FILE	ZATION ON	YES		120
Sample ID		. = 3		
Tested	05/08/24			
Tester	JRIO			
	05/08/24			
PH OF PRODUCT/SOLU		7.200	7.150 - 7.250	7.202
	-	PH		
Sample ID				
Tested	05/08/24			
Tester	JRIO			
	05/08/24			
FILTERED THROUGH 0.2 UM FILTI		PASS		PASS
	_	PASS		
Sample ID				
Tested	05/08/24			
Tester	JRIO			
	05/08/24			
APPEARANCE AND COL		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	05/08/24			
Tester	JRIO			
	05/08/24			
NSTRUMENTS USED D		INSTRUMENT		M103,T190
PREPARATION				
Sample ID				
Tested	05/08/24			
Tester	JRIO			
	05/08/24			
EXPIRATION DATE		MM/DD/YY		05/30/25
		MM/DD/YYYY		
Sample ID				
Tested	05/08/24			
Tester	JRIO			
Recorded	05/08/24			
PREPARED TO STANDA	RD	PASS		PASS
METHODS		YES		
Sample ID				
Tested	05/08/24			
Tester	JRIO			
Recorded	05/08/24			
SHELF LIFE OF PRODU	СТ	1 YEAR		1 YEAR
		TIME		
Sample ID				
Tested	05/08/24			
Tester	JRIO			
Described	05/00/04			

Page: 2

Date: 05/08/24 at 1:32 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Lot # 4050824 Made 05/08/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**