



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 05/06/24			
Tester JENC			
Recorded 05/06/24			
PH OF PRODUCT/SOLUTION	7.000	6.000 - 12.500	10.290
	PH		
Sample ID			
Tested 05/06/24			
Tester JENC			
Recorded 05/06/24			
Actual Normality	0.1000	0.0995 - 0.1005	0.1003
	EQ/L		
Sample ID			
Tested 05/06/24			
Tester JENC			
Recorded 05/06/24			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP10		
Sample ID			
Tested 05/06/24			
Tester JENC			
Recorded 05/06/24			
QCSP35:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP35		
Sample ID			
Tested 05/06/24			
Tester JENC			
Recorded 05/06/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 05/06/24			
Tester JENC			
Recorded 05/06/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,M112,T209,M55,M91,BID0 2259,TID
Sample ID			
Tested 05/06/24			
Tester JENC			
Recorded 05/06/24			
DATE OF PREPARATION	MM/DD/YYYY		05/06/2024
Sample ID			
Tested 05/06/24			
Tester JENC			
Recorded 05/06/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		11/30/24
Sample ID			
Tested	05/06/24		
Tester	JENC		
Recorded	05/06/24		
SHELF LIFE	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested	05/06/24		
Tester	JENC		
Recorded	05/06/24		
TRACEABLE TO NIST SRM 136	YES YES		YES
Sample ID			
Tested	05/06/24		
Tester	JENC		
Recorded	05/06/24		
Actual Density	1.01127 GM/ML	1.01080 - 1.01173	1.01123
Sample ID			
Tested	05/06/24		
Tester	JENC		
Recorded	05/06/24		
Lot #	4050360		
Made	05/06/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager