

Page: 1

Date: 05/07/24 at 1:54 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: UC23			
	UM ACETATE 0.1 NORMAL IN		
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION OF			YES
FILE	YES		
Sample ID			
Tested 05/07/24			
Tester AFRE			
Recorded 05/07/24			
Actual Normality	0.1000	0.0991 - 0.1009	0.0998
	EQ/L		
Sample ID			
Tested 05/07/24			
Tester AFRE			
Recorded 05/07/24			
Actual Density	1.05347	1.05260 - 1.05433	1.05370
	GM/ML		
Sample ID			
Tested 05/07/24			
Tester AFRE			
Recorded 05/07/24			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID	-		
Tested 05/07/24			
Tester AFRE			
Recorded 05/07/24			
INSTRUMENTS USED DURING	INSTRUMENT		M103,GC,M55,VF12-3,BID0048
PREPARATION	INSTITUMENT		8
			O
Sample ID			
Tested 05/07/24			
Tester AFRE			
Recorded 05/07/24	O MONITUO		2 MONTHO
SHELF LIFE OF PRODUCT	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested 05/07/24			
Tester AFRE			
Recorded 05/07/24			· ·
EXPIRATION DATE	MM/DD/YYYY		11/30/2024
	MM/DD/YYYY		
Sample ID			
Tested 05/07/24			
Tester AFRE			
Recorded 05/07/24			
Lot # 4041649			
Made 05/07/24			
IVIAUE 00/01/21			



Page: 2

Date: 05/07/24 at 1:54 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager