



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: H1953-4L | | | |
|--|-------------------|-------|-----------------|
| HEXANES (AS n-hexane) >95% HPLC (SUITABLE FOR O&G) (4 LITERS) | | | |
| Test | Target/UOM | Range | Result |
| ASSAY OF MATERIAL | MIN 95% | | 97.6% |
| | % | | |
| Sample ID | | | |
| Tested | 01/19/23 | | |
| Tester | TBUR | | |
| Recorded | 01/19/23 | | |
| BOILING RANGE-ENTIRE | 66.1 - 69.4 DEG C | | 66.6-68.9 DEG C |
| | DEG | | |
| Sample ID | | | |
| Tested | 01/19/23 | | |
| Tester | TBUR | | |
| Recorded | 01/19/23 | | |
| COLOR | MAX 10 APHA | | <10 APHA |
| | PT-CO | | |
| Sample ID | | | |
| Tested | 01/19/23 | | |
| Tester | TBUR | | |
| Recorded | 01/19/23 | | |
| FLUORESCENCE (AS QUININE BASE) | MAX 250 PPT | | 85 PPT |
| | FLUORESCEN | | |
| Sample ID | | | |
| Tested | 01/19/23 | | |
| Tester | TBUR | | |
| Recorded | 01/19/23 | | |
| FORM | CLEAR LIQUID | | CLEAR LIQUID |
| | FORM | | |
| Sample ID | | | |
| Tested | 01/19/23 | | |
| Tester | TBUR | | |
| Recorded | 01/19/23 | | |
| INFRARED | CONFORMS | | CONFORMS |
| | INFRARED | | |
| Sample ID | | | |
| Tested | 01/19/23 | | |
| Tester | TBUR | | |
| Recorded | 01/19/23 | | |
| RESIDUE AFTER EVAPORATION | MAX 1 PPM | | <0.5 PPM |
| Sample ID | | | |
| Tested | 01/19/23 | | |
| Tester | TBUR | | |
| Recorded | 01/19/23 | | |
| UV CUTOFF | MAX 195 NM | | 189.4 NM |
| | NM | | |
| Sample ID | | | |
| Tested | 01/19/23 | | |
| Tester | TBUR | | |
| Recorded | 01/19/23 | | |



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| | | |
|----------------------|-----------|---------|
| WATER CONTENT | MAX 0.01% | 0.0020% |
| | % | |
| Sample ID | | |
| Tested | 01/19/23 | |
| Tester | TBUR | |
| Recorded | 01/19/23 | |
| Lot # | 301327 | |
| Made | 01/13/23 | |

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager