



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: I7604-4L			
ISOPROPYL ALCOHOL HPLC GRADE (PX1834-1)			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	MIN 99.5%		>99.99%
	%		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
WATER CONTENT	MAX 0.01%		0.0013%
	%		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
UV CUTOFF	204 NM		202.4 NM
	NM		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
REFRACTIVE INDEX	1.3700	0.0000 - 1.3800	1.3771
	REFRACTIVE INDE		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
ELECTRON CAPTURE	PASSES TEST		PASSES TEST
	ELECTRON C		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
RESIDUE AFTER EVAPORATION	MAX 1 PPM		<0.5 PPM
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
MANUFACTURED DATE OF MATERIAL	MM/DD/YYYY		04/20/2023
	MM/DD/YYYY		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
Lot #	312801		
Made	05/08/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 05/12/23 at 9:44 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager