

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	I7600			
		PYL ALCOHOL REAGENT ACS		
Test		Target/UOM	Range	Result
ACETONE %		<=0.002	.	0.002%
		%		
Sample ID				
Tested	05/19/23			
Tester	TBUR			
Recorded	05/19/23			
ASSAY OF MATERIAL		>= 99.5%		99.9%
		%		
Sample ID				
Tested	05/19/23			
Tester	TBUR			
Recorded	05/19/23			
COLOR		MAX 10 APHA		<5 APHA
		PT-CO		
Sample ID				
Tested	05/19/23			
Tester	TBUR			
Recorded	05/19/23			
SOLUBILITY		PASSES TEST		PASSES TEST
		PASS		
Sample ID				
Tested	05/19/23			
Tester	TBUR			
Recorded	05/19/23			0.01007
WATER CONTENT		MAX 0.2%		0.01367
		%		
Sample ID				
Tested	05/19/23			
Tester	TBUR			
Recorded	05/19/23	0.0010/		0.000000/
RESIDUE AFTER EVAP	ORATION	< = 0.001%		0.00003%
Sample ID				
Tested	05/19/23			
Tester	TBUR			
Recorded	05/19/23			
TITRABLE ACID		MAX 0.0001 MEQ/G		0.00004 MEQ/G
		T ACID		
Sample ID	0.5/10/00			
Tested	05/19/23			
Tester	TBUR			
Recorded TITRABLE BASE	05/19/23	MAX 0.0001 MEQ/G		0.00004 MEQ/G
		T BASE		0.00004 MEQ/G
Sample ID				
Tested	05/19/23			
Tester	TBUR			
Recorded	05/19/23			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 **Certificate of Analysis / QC Results**

Customer PO # :

EXPIRATION DATE		MM/DD/YY MM/DD/YYYY	09/30/27
Sample ID			
Tested	05/19/23		
Tester	TBUR		
Recorded	05/19/23		
SHELF LIFE OF PRODU		SHELF LIFE	5 YEARS
	-	TIME	
Sample ID			
Tested	05/19/23		
Tester	TBUR		
Recorded	05/19/23		
MANUFACTURE DATE		MM/DD/YY	09/13/22
		MM/DD/YYYY	
Sample ID			
Tested	05/19/23		
Tester	TBUR		
Recorded	05/19/23		
Lot # 31380	5		
Made 05/18			

hichard Sto

Quality Manager