



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: S2599			
SODIUM CARBONATE RGNT ACS ANHYDROUS			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	>= 99.5%		99.94%
	%		
Sample ID			
Tested	07/21/23		
Tester	TBUR		
Recorded	07/21/23		
INSOLUBLE MATTER	MAX 0.01%		0.008%
	%		
Sample ID			
Tested	07/21/23		
Tester	TBUR		
Recorded	07/21/23		
LOSS ON HEATING AT 285 DEG C	MAX 1.0%		0.42%
	%		
Sample ID			
Tested	07/21/23		
Tester	TBUR		
Recorded	07/21/23		
CHLORIDE CONTENT	MAX 0.001%		<0.001%
	PPM		
Sample ID			
Tested	07/21/23		
Tester	TBUR		
Recorded	07/21/23		
PHOSPHATE CONTENT	MAX 0.001%		<0.001%
	PPM		
Sample ID			
Tested	07/21/23		
Tester	TBUR		
Recorded	07/21/23		
SiO2	MAX 0.005%		<0.005%
	PPB		
Sample ID			
Tested	07/21/23		
Tester	TBUR		
Recorded	07/21/23		
SULFUR COMPOUNDS (AS S)	MAX 0.003%		<0.003%
	%		
Sample ID			
Tested	07/21/23		
Tester	TBUR		
Recorded	07/21/23		
HEAVY METAL (AS PB)	MAX 5 PPM		<5 PPM
	PPM		
Sample ID			
Tested	07/21/23		
Tester	TBUR		
Recorded	07/21/23		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

IRON	MAX 5 PPM PPM	<5 PPM
Sample ID		
Tested	07/21/23	
Tester	TBUR	
Recorded	07/21/23	
CALCIUM CONTENT	MAX 0.03% PPM	<0.03%
Sample ID		
Tested	07/21/23	
Tester	TBUR	
Recorded	07/21/23	
MAGNESIUM (MG) CONTENT	MAX 0.005% PPM	<0.005%
Sample ID		
Tested	07/21/23	
Tester	TBUR	
Recorded	07/21/23	
POTASSIUM CONTENT	MAX 0.005% PPM	<0.005%
Sample ID		
Tested	07/21/23	
Tester	TBUR	
Recorded	07/21/23	
MANUFACTURE DATE	MM/DD/YYYY MM/DD/YYYY	02/22/22
Sample ID		
Tested	07/21/23	
Tester	TBUR	
Recorded	07/21/23	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	02/21/25
Sample ID		
Tested	07/21/23	
Tester	TBUR	
Recorded	07/21/23	
Lot #	320108	
Made	07/20/23	

Quality Manager