

Page: 1

Date: 05/14/24 at 1:33 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produ				
	CONDUC	IVITY STANDARD 15.0 uMHOS		
Test		Target/UOM	Range	Result
PREPARED TO FOR	MULATION ON	YES		YES
FILE		YES		
Sample	ID			
Tested	05/14/24			
Tester	JRIO			
Recorde				
CONDUCTIVITY @2!	5 C	15	14 - 16	15
		US/CM		
Sample	ID			
Tested	05/14/24			
Tester	JRIO			
Recorde	ed 05/14/24			
QCSP30:CONDUCTI	VITY	YES		YES
STD&MEASUREMENT		QCSP30		
Sample	ID			
Tested	05/14/24			
Tester	JRIO			
Recorde	ed 05/14/24			
TRACEABLE TO N.I.	S.T. SRM 3191	YES		YES
		YES		
Sample	ID			
Tested	05/14/24			
Tester	JRIO			
Recorde				
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample	ID			
Tested	05/14/24			
Tester	JRIO			
Recorde				
INSTRUMENTS USE		INSTRUMENT		T234,M92
PREPARATION				1207,1402
Sample	ID			
Tested	05/14/24			
Tester	05/14/24 JRIO			
Recorde				
EXPIRATION DATE	JG 03/14/24	MM/DD/YYYY		11/30/2024
EXPIRATION DATE		MM/DD/YYYY		11/30/2024
Comple	ID	IVIIVI/DD/1111		
Sample				
Tested	05/14/24 JRIO			
Tester Recorde				
Lot # 405				
<b>Made</b> 05/	14/24			

Page: 2

Date: 05/14/24 at 1:33 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**