



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: VEO003F			
Calibration Standard Stock Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 05/15/24			
Tester MMAT			
Recorded 05/15/24			
Calculated ppm E2015	8100 PPM	7600 - 8600	8119
Sample ID			
Tested 05/15/24			
Tester MMAT			
Recorded 05/15/24			
Calculated ppm G7510	3875 PPM	3375 - 4375	3870
Sample ID			
Tested 05/15/24			
Tester MMAT			
Recorded 05/15/24			
Calculated ppm D4690	12300 PPM	11800 - 12800	12317
Sample ID			
Tested 05/15/24			
Tester MMAT			
Recorded 05/15/24			
Calculated ppm D4691	12450 PPM	11950 - 12950	12436
Sample ID			
Tested 05/15/24			
Tester MMAT			
Recorded 05/15/24			
Calculated ppm C3780	14350 PPM	13850 - 14850	14330
Sample ID			
Tested 05/15/24			
Tester MMAT			
Recorded 05/15/24			
APPEARANCE AND COLOR	PASS APPEARANCE		PASS
Sample ID			
Tested 05/15/24			
Tester MMAT			
Recorded 05/15/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M115
Sample ID			
Tested 05/15/24			
Tester MMAT			
Recorded 05/15/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 05/15/24 at 4:58 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE

MM/DD/YYYY

05/30/2025

MM/DD/YYYY

Sample ID

Tested 05/15/24

Tester MMAT

Recorded 05/15/24

Lot # 4040921

Made 05/15/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager