



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: A1520			
ACETONITRILE HPLC GRADE			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	MIN 99.5%		99.99%
	%		
Sample ID			
Tested	02/08/24		
Tester	TBUR		
Recorded	02/08/24		
WATER CONTENT	MAX 0.02%		0.002%
	%		
Sample ID			
Tested	02/08/24		
Tester	TBUR		
Recorded	02/08/24		
EXPIRATION DATE	MM/DD/YY		12/30/2028
	MM/DD/YYYY		
Sample ID			
Tested	02/08/24		
Tester	TBUR		
Recorded	02/08/24		
SHELF LIFE OF PRODUCT	SHELF LIFE		5 YEARS
	TIME		
Sample ID			
Tested	02/08/24		
Tester	TBUR		
Recorded	02/08/24		
MANUFACTURE DATE	MM/DD/YY		12/13/2023
	MM/DD/YYYY		
Sample ID			
Tested	02/08/24		
Tester	TBUR		
Recorded	02/08/24		
Lot #	403012		
Made	01/30/24		

Quality Manager