



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: I7604			
ISOPROPYL ALCOHOL HPLC GRADE			
Test	Target/UOM	Range	Result
UV CUTOFF	204 NM NM		202.1 NM
Sample ID			
Tested	04/03/24		
Tester	TBUR		
Recorded	04/03/24		
REFRACTIVE INDEX	1.3760 REFRACTIVE INDE	1.3760 - 1.3800	1.3771
Sample ID			
Tested	04/03/24		
Tester	TBUR		
Recorded	04/03/24		
ASSAY OF MATERIAL	MIN 99.5% %		>99.99%
Sample ID			
Tested	04/03/24		
Tester	TBUR		
Recorded	04/03/24		
ELECTRON CAPTURE GC	PASSES TEST ELECTRONCA		PASSES TEST
Sample ID			
Tested	04/03/24		
Tester	TBUR		
Recorded	04/03/24		
WATER CONTENT	MAX 0.05% %		0.0047%
Sample ID			
Tested	04/03/24		
Tester	TBUR		
Recorded	04/03/24		
RESIDUE AFTER EVAPORATION	MAX 1 PPM RESEVAP		<0.5 PPM
Sample ID			
Tested	04/03/24		
Tester	TBUR		
Recorded	04/03/24		
MANUFACTURE DATE	MM/DD/YY MM/DD/YYYY		02/08/2024
Sample ID			
Tested	04/03/24		
Tester	TBUR		
Recorded	04/03/24		
Lot #	409216		
Made	04/01/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 04/03/24 at 2:25 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager