Page: 1

Date: 05/17/24 at 10:22 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

- 9	ed Product:		STANDARD 100 PPM AS NI	H3 IN 0.1N H2SO4	
Test			Target/UOM	Range	Result
	ED TO FORMU	LATION ON	YES	3	YES
FILE			YES		
	Sample ID				
	Tested	05/17/24			
	Tester	AROS			
	Recorded	05/17/24			
CERTIFIE	D PPM NH3		100.0	96.0 - 104.0	100.0
			PPM		
	Sample ID				
	Tested	05/17/24			
	Tester	AROS			
	Recorded	05/17/24			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	05/17/24			
	Tester	AROS			
	Recorded	05/17/24			
NSTRUM	IENTS USED D	URING	INSTRUMENT		P151,VF500-X42
PREPARA	ATION				
	Sample ID				
	Tested	05/17/24			
	Tester	AROS			
	Recorded	05/17/24			
EXPIRATION DATE			MM/DD/YYYY		11/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	05/17/24			
	Tester	AROS			
	Recorded	05/17/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Prichard St.