Page: 1

Date: 03/19/24 at 10:31 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	Ou Ou	3.0πC 1 O π .	
Packaged Product:	BAS088		
· ·	REDOX CALIBRATION STD. 200m	nV +/10 GLI#25M2A1001-115 (500ML)	
Test	Target/UOM	Range	Result
MILLIVOLT READING	195 TO 205 MV	<u> </u>	201.0 MV
	MV		
Sample ID			
Tested	03/19/24		
Tester	TBUR		
Recorded	03/19/24		
EXPIRATION DATE	MM/DD/YY		02/28/2025
	MM/DD/YYYY		
Sample ID			
Tested	03/19/24		
Tester	TBUR		
Recorded	03/19/24		
MANUFACTURE DATE	MM/DD/YY		02/19/2024
	MM/DD/YYYY		
Sample ID			
Tested	03/19/24		
Tester	TBUR		
Recorded	03/19/24		
SHELF LIFE OF PRODUC	T 1 YEAR		1 YEAR
	TIME		
Sample ID			
Tested	03/19/24		
Tester	TBUR		
Recorded	03/19/24		
Lot # 407809			
Made 03/18/	24		

Quality Manager