

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	1660			
•		CHLORIDE 0.01 MOLAR		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	05/16/24			
Tester	AFRE			
Recorded	05/16/24			
Actual Molarity		0.0100	0.0097 - 0.0103	0.0100
		MOL/L		
Sample ID				
Tested	05/16/24			
Tester	AFRE			
Recorded	05/16/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	05/16/24			
Tester	AFRE			
Recorded	05/16/24			
Actual Density		0.99916	0.99903 - 0.99929	0.99914
		GM/ML		
Sample ID				
Tested	05/17/24			
Tester	AFRE			
Recorded	05/17/24			
INSTRUMENTS USED D	URING	INSTRUMENT		M91,VF2-A7,BID-03828,PID-54
PREPARATION				
Sample ID				
Tested	05/17/24			
Tester	AFRE			
Recorded	05/17/24			
N.I.S.T. Traceable to SRM 915		YES		YES
		YES		
Sample ID				
Tested	05/17/24			
Tester	AFRE			
Recorded	05/17/24			
PREPARED WITH REAC		YES		YES
GRADE MATERIAL		ACS GRADE		120
Sample ID				
Tested	05/17/24			
Tester	AFRE			
Recorded	05/17/24			
EXPIRATION DATE	00/17/24	MM/DD/YY		05/30/25
		MM/DD/YYYY		03/30/23
Camale ID				
Sample ID	05/17/04			
Tested	05/17/24			
Tester	AFRE			
Recorded	05/17/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4051403
Made	05/17/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard St

Quality Manager