05/14/24

Made

Page: 1

Date: 05/14/24 at 9:57 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Pr				
	Phospha	te Buffer, pH 5.2		
Test		Target/UOM	Range	Result
PREPARED TO	FORMULATION ON	YES		YES
FILE		YES		
Sai	mple ID			
Tes	sted 05/14/24			
Tes	ster BFRA			
Re	corded 05/14/24			
PH OF PRODUCT/SOLUTION		5.200	5.100 - 5.300	5.122
		PH		
Sai	mple ID			
Tes	sted 05/14/24			
Tes	ster BFRA			
Re	corded 05/14/24			
APPEARANCE AND COLOR		PASS		PASS
		APPEARANCE		
Sai	mple ID			
	sted 05/14/24			
Tes	ster BFRA			
Re	corded 05/14/24			
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF1-Z44
PREPARATION				
Sai	mple ID			
	sted 05/14/24			
Tes	ster BFRA			
Re	corded 05/14/24			
EXPIRATION DATE		MM/DD/YYYY		05/30/2025
		MM/DD/YYYY		
Sai	mple ID			
	sted 05/14/24			
	ster BFRA			
Re	corded 05/14/24			
Lot #	4051372			
	05 /14 /04			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St