



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: S2675			
SODIUM CHLORIDE REAGENT ACS CRYSTAL			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	>= 99.0%		99.32%
	%		
Sample ID			
Tested	05/13/24		
Tester	TBUR		
Recorded	05/13/24		
PH OF A 5% SOLUTION	5.0 - 9.0 AT 25 DEG C		6.30
	PH		
Sample ID			
Tested	05/13/24		
Tester	TBUR		
Recorded	05/13/24		
INSOLUBLE MATTER	MAX 0.005%		0.004%
	%		
Sample ID			
Tested	05/13/24		
Tester	TBUR		
Recorded	05/13/24		
IODIDE	MAX 0.002%		<0.002%
	PPM		
Sample ID			
Tested	05/13/24		
Tester	TBUR		
Recorded	05/13/24		
CHLORATE AND NITRATE (AS NO3)	MAX 0.003%		<0.003%
	CHLORATE		
Sample ID			
Tested	05/13/24		
Tester	TBUR		
Recorded	05/13/24		
PHOSPHATE CONTENT	MAX 5 PPM		<5 PPM
	PPM		
Sample ID			
Tested	05/13/24		
Tester	TBUR		
Recorded	05/13/24		
BARIUM	PASSES TEST		PASSES TEST
	BARIUM		
Sample ID			
Tested	05/13/24		
Tester	TBUR		
Recorded	05/13/24		
HEAVY METAL (AS PB)	MAX 5 PPM		<5 PPM
	PPM		
Sample ID			
Tested	05/13/24		
Tester	TBUR		
Recorded	05/13/24		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

IRON	MAX 2 PPM PPM	<2 PPM
Sample ID		
Tested	05/13/24	
Tester	TBUR	
Recorded	05/13/24	
CALCIUM CONTENT	MAX 0.002% PPM	<0.002%
Sample ID		
Tested	05/13/24	
Tester	TBUR	
Recorded	05/13/24	
MAGNESIUM (MG) CONTENT	MAX 0.001% PPM	<0.001%
Sample ID		
Tested	05/13/24	
Tester	TBUR	
Recorded	05/13/24	
POTASSIUM CONTENT	MAX 0.005% PPM	<0.005%
Sample ID		
Tested	05/13/24	
Tester	TBUR	
Recorded	05/13/24	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	01/30/2027
Sample ID		
Tested	05/13/24	
Tester	TBUR	
Recorded	05/13/24	
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	3 YEARS
Sample ID		
Tested	05/13/24	
Tester	TBUR	
Recorded	05/13/24	
Lot #	412905	
Made	05/08/24	

Quality Manager