



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: N3805-2.5L			
NITRIC ACID 68-70% REAGENT ACS (2.5L) (NX0409-2)			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR	TO PASS TEST APPEARANCE		PASS
Sample ID			
Tested	05/16/23		
Tester	TBUR		
Recorded	05/16/23		
ASSAY OF MATERIAL	68.0% - 70.0% %		69.55%
Sample ID			
Tested	05/16/23		
Tester	TBUR		
Recorded	05/16/23		
COLOR	MAX 10 APHA PT-CO		<10 APHA
Sample ID			
Tested	05/16/23		
Tester	TBUR		
Recorded	05/16/23		
RESIDUE OF IGNITION	MAX 5 PPM RESIDUE IG		<5 PPM
Sample ID			
Tested	05/16/23		
Tester	TBUR		
Recorded	05/16/23		
CHLORIDE CONTENT	MAX 0.5 PPM PPM		<0.5 PPM
Sample ID			
Tested	05/16/23		
Tester	TBUR		
Recorded	05/16/23		
PHOSPHATE CONTENT	MAX 0.2 PPM PPM		<0.2 PPM
Sample ID			
Tested	05/16/23		
Tester	TBUR		
Recorded	05/16/23		
SULFATE (SO4) CONTENT	MAX 1 PPM MAX 60 MG/KG		<1 PPM
Sample ID			
Tested	05/16/23		
Tester	TBUR		
Recorded	05/16/23		
ALUMINUM (AL) CONTENT	MAX 0.2 PPM		0.007 PPM
Sample ID			
Tested	05/16/23		
Tester	TBUR		
Recorded	05/16/23		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

ARSENIC CONTENT	MAX 0.01 PPM	<0.01 PPM
	ARSENIC	
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
CALCIUM CONTENT	MAX 0.2 PPM	0.031 PPM
	PPM	
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
CHROMIUM CONTENT	MAX 0.1 PPM	0.002 PPM
	PPM	
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
COPPER (CU) CONTENT	MAX 0.05 PPM	0.001 PPM
	COPPER	
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
GOLD (AU)	MAX 0.3 PPM	0.001 PPM
	GOLD	
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
HEAVY METAL (AS PB)	MAX 0.2 PPM	<0.2 PPM
	PPM	
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
IRON	MAX 0.2 PPM	0.010 PPM
	PPM	
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
LEAD (PB) CONTENT	MAX 0.1 PPM	0.001 PPM
	MG/L	
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

MAGNESIUM (MG) CONTENT	MAX 0.3 PPM PPM	0.007 PPM
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
MANGANESE (MN) CONTENT	MAX 0.2 PPM %	0.001 PPM
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
MERCURY (HG) CONTENT	MAX 5 PPB PPM	<5 PPB
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
NICKEL (NI) CONTENT	MAX 0.05 PPM NICKEL	0.001 PPM
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
POTASSIUM CONTENT	MAX 0.5 PPM PPM	0.010 PPM
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
SODIUM CONTENT	MAX 0.5 PPM %	0.024 PPM
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	
TIN (SN) CONTENT	MAX 0.3 PPM PPM	0.006 PPM
Sample ID		
Tested	05/16/23	
Tester	TBUR	
Recorded	05/16/23	

Lot # 313512

Made 05/15/23



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 4

Date: 05/16/23 at 10:15 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager