



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: BAS088A			
REDOX CALIBRATION STD. 200mV +/-10 GLI#25M2A1001-123 (3.5L)			
Test	Target/UOM	Range	Result
MILLIVOLT READING	195 TO 205 MV MV		199.0 MV
Sample ID			
Tested	11/17/23		
Tester	TBUR		
Recorded	11/17/23		
EXPIRATION DATE	MM/YYYY MM/DD/YYYY		10/31/2024
Sample ID			
Tested	11/17/23		
Tester	TBUR		
Recorded	11/17/23		
MANUFACTURE DATE	MM/DD/YY MM/DD/YYYY		10/23/2023
Sample ID			
Tested	11/17/23		
Tester	TBUR		
Recorded	11/17/23		
SHELF LIFE OF PRODUCT	1 YEAR TIME		1 YEAR
Sample ID			
Tested	11/17/23		
Tester	TBUR		
Recorded	11/17/23		
Lot #	331805		
Made	11/14/23		

Quality Manager