



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: N3603			
NINHYDRIN (EMD BRAND ONLY)			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	> 99.0%		99.6%
	%		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
IDENTITY	CONFORMS		CONFORMS
	PASS		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
PH OF PRODUCT/SOLUTION	4.900	4.600 - 5.600	5.300
	PH		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
SULFATED ASH	MAX 0.1%		<0.1%
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
EXPIRATION DATE	MM/DD/YY		01/24/26
	MM/DD/YYYY		
Sample ID			
Tested	05/12/23		
Tester	TBUR		
Recorded	05/12/23		
Lot #	312408		
Made	05/04/23		

Quality Manager