



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: A1520			
ACETONITRILE HPLC GRADE			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	MIN 99.5%		99.9%
	%		
Sample ID			
Tested	07/11/23		
Tester	TBUR		
Recorded	07/11/23		
WATER CONTENT	MAX 0.02%		<0.01%
	%		
Sample ID			
Tested	07/11/23		
Tester	TBUR		
Recorded	07/11/23		
EXPIRATION DATE	MM/DD/YY		05/30/28
	MM/DD/YYYY		
Sample ID			
Tested	07/11/23		
Tester	TBUR		
Recorded	07/11/23		
SHELF LIFE OF PRODUCT	SHELF LIFE		5 YEARS
	TIME		
Sample ID			
Tested	07/11/23		
Tester	TBUR		
Recorded	07/11/23		
MANUFACTURE DATE	MM/DD/YY		05/16/23
	MM/DD/YYYY		
Sample ID			
Tested	07/11/23		
Tester	TBUR		
Recorded	07/11/23		
Lot #	319104		
Made	07/10/23		

Quality Manager