



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: M2127			
MERCURIC CHLORIDE REAGENT ACS			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	>= 99.5%		99.69%
	%		
Sample ID			
Tested 01/22/24			
Tester TBUR			
Recorded 01/22/24			
SOLUTION IN ETHYL ETHER	PASSES TEST		PASSES TEST
	PASS		
Sample ID			
Tested 01/22/24			
Tester TBUR			
Recorded 01/22/24			
IRON	MAX 0.002%		<0.002%
	PPM		
Sample ID			
Tested 01/22/24			
Tester TBUR			
Recorded 01/22/24			
RESIDUE AFTER REDUCTION	MAX 0.02%		0.006%
Sample ID			
Tested 01/22/24			
Tester TBUR			
Recorded 01/22/24			
Lot # 401923			
Made 01/19/24			

Quality Manager