



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ERL265 BROMINE SOLUTION (1L)			
Lot # 3011067			
Chemical Product: ERL265F BROMINE SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE		YES YES	YES
Sample ID			
Tested	01/13/23		
Tester	MHUS		
Recorded	01/13/23		
APPEARANCE AND COLOR		CLEAR ORANGE LIQUID APPEARANCE	PASS
Sample ID			
Tested	01/13/23		
Tester	MHUS		
Recorded	01/13/23		
INSTRUMENTS USED DURING PREPARATION		INSTRUMENT	M100,GC,VF5-L8
Sample ID			
Tested	01/13/23		
Tester	MHUS		
Recorded	01/13/23		
SHELF LIFE OF PRODUCT		6 WEEKS TIME	6 WEEKS
Sample ID			
Tested	01/13/23		
Tester	MHUS		
Recorded	01/13/23		
EXPIRATION DATE		MM/DD/YYYY MM/DD/YYYY	02/24/23
Sample ID			
Tested	01/13/23		
Tester	MHUS		
Recorded	01/13/23		
Made 01/13/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 02/09/23 at 4:29 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager