



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> I7604-4L			
ISOPROPYL ALCOHOL HPLC GRADE (PX1834-1)			
Test	Target/UOM	Range	Result
<b>ASSAY OF MATERIAL</b>	MIN 99.5%		>99.99%
	%		
Sample ID			
Tested	08/08/23		
Tester	TBUR		
Recorded	08/08/23		
<b>WATER CONTENT</b>	MAX 0.01%		0.0040%
	%		
Sample ID			
Tested	08/08/23		
Tester	TBUR		
Recorded	08/08/23		
<b>UV CUTOFF</b>	204 NM		202.2 NM
	NM		
Sample ID			
Tested	08/08/23		
Tester	TBUR		
Recorded	08/08/23		
<b>REFRACTIVE INDEX</b>	1.3700	0.0000 - 1.3800	1.3771
	REFRACTIVE INDE		
Sample ID			
Tested	08/08/23		
Tester	TBUR		
Recorded	08/08/23		
<b>ELECTRON CAPTURE</b>	PASSES TEST		PASSES TEST
	ELECTRON C		
Sample ID			
Tested	08/08/23		
Tester	TBUR		
Recorded	08/08/23		
<b>RESIDUE AFTER EVAPORATION</b>	MAX 1 PPM		<0.5 PPM
Sample ID			
Tested	08/08/23		
Tester	TBUR		
Recorded	08/08/23		
<b>MANUFACTURED DATE OF MATERIAL</b>	MM/DD/YYYY		07/18/2023
	MM/DD/YYYY		
Sample ID			
Tested	08/08/23		
Tester	TBUR		
Recorded	08/08/23		
<b>Lot #</b>	321901		
<b>Made</b>	08/07/23		



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 08/08/23 at 12:35 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**