



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: A1520-4L			
ACETONITRILE HPLC GRADE (4L) (AX0142-1)			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	MIN 99.9%		>99.99%
	%		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
COLOR	MAX 10		<10 APHA
	PT-CO		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
RESIDUE AFTER EVAPORATION	MAX 1 PPM		<0.5 PPM
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
FILTERED THROUGH 0.2 UM FILTER	PASSES TEST		PASSES TEST
	YES		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
FLUORESCENCE (AS QUININE BASE)	MAX 300 PPT		94 PPT
	FLUORESCEN		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
FORM	CLEAR LIQUID		CLEAR LIQUID
	FORM		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
REFRACTIVE INDEX	1.3434	1.3430 - 1.3450	1.3438
	REFRACTIVE INDE		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
TITRATABLE ACID	0.50	0.00 - 8.00	0.32
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Titration base	0.10	0.00 - 0.10	0.08
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
UV CUTOFF	MAX 189 NM		187.4 NM
	NM		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
WATER CONTENT	MAX 0.02%		0.0021%
	%		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
IDENTITY	CONFORMS		CONFORMS
	PASS		
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
DATE OF MANUFACTURE	DD/MM/YYYY		09/06/2023
Sample ID			
Tested	09/28/23		
Tester	TBUR		
Recorded	09/28/23		
Lot #	326106		
Made	09/18/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager