

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	LY127F			
	CYANIDE STANDA	ARD 10.0 PPM AQUEOUS		
Test	Tar	get/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	05/17/24			
Tester	MMAT			
Recorded	05/17/24			
APPEARANCE AND COL	.OR CLE	AR LIQUID		PASS
	APF	PEARANCE		
Sample ID				
Tested	05/17/24			
Tester	MMAT			
Recorded	05/17/24			
INSTRUMENTS USED DU	URING INS	TRUMENT		P98,VF100-K4
PREPARATION				
Sample ID				
Tested	05/17/24			
Tester	MMAT			
	05/17/24			
Actual ppm Cyanide		10.00	9.80 - 10.20	9.97
		PPM		
Sample ID				
Tested	05/17/24			
Tester	MMAT			
Recorded	05/17/24			
EXPIRATION DATE	М	M/DD/YY		07/22/24
	MM	/DD/YYYY		
Sample ID				
Tested	05/17/24			
	MMAT			
Recorded	05/17/24			
Lot # 4051782				
Made 05/17,	/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA