



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: BP068
 ISO-OCTANE ASTM PRF (T8259) (12.9KGS)

Lot # 331006

Chemical Product: T8259
 TRIMETHYLPENTANE (ISO OCTANE) ASTM PRIMARY REFERENCE FUEL

Test	Target/UOM	Range	Result
LEAD (PB) CONTENT	MG/L MG/L		<0.5 MG/L
Sample ID			
Tested	11/09/23		
Tester	TBUR		
Recorded	11/09/23		
ASSAY N-HEPTANE	MAX <= 0.10 LV% % W/W		<0.1%
Sample ID			
Tested	11/09/23		
Tester	TBUR		
Recorded	11/09/23		
ASSAY OF MATERIAL	> = 99.75% %		99.81%
Sample ID			
Tested	11/09/23		
Tester	TBUR		
Recorded	11/09/23		
Sulfur compounds (as S)	0.005% MAX % W/W		<0.005%
Sample ID			
Tested	11/09/23		
Tester	TBUR		
Recorded	11/09/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		10/19/2025
Sample ID			
Tested	11/09/23		
Tester	TBUR		
Recorded	11/09/23		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		2 YEARS
Sample ID			
Tested	11/09/23		
Tester	TBUR		
Recorded	11/09/23		
MANUFACTURE DATE	MM/DD/YY MM/DD/YYYY		10/20/2023
Sample ID			
Tested	01/15/24		
Tester	TBUR		
Recorded	01/15/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 01/15/24 at 12:29 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Made 11/06/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager