



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS107F			
Conductivity Standard 74.0 umhos @25°C			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 05/21/24			
Tester JRIO			
Recorded 05/21/24			
CONDUCTIVITY @25 C.	74 US/CM	73 - 75	74
Sample ID			
Tested 05/21/24			
Tester JRIO			
Recorded 05/21/24			
QCSP30: CONDUCTIVITY STD&MEASUREMENT	YES QCSP30		YES
Sample ID			
Tested 05/21/24			
Tester JRIO			
Recorded 05/21/24			
TRACEABLE TO N.I.S.T. SRM 3191	YES YES		YES
Sample ID			
Tested 05/21/24			
Tester JRIO			
Recorded 05/21/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 05/21/24			
Tester JRIO			
Recorded 05/21/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T234,M92
Sample ID			
Tested 05/21/24			
Tester JRIO			
Recorded 05/21/24			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		05/30/2025
Sample ID			
Tested 05/21/24			
Tester JRIO			
Recorded 05/21/24			
Lot #	4052178		
Made	05/21/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 05/21/24 at 2:42 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager