

Page: 1

Date: 05/20/24 at 2:36 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## **Certificate of Analysis / QC Results**

Customer PO #:

Packaged Product	<b>:</b> 3860			
		E STANDARD 10 PPM (1ML	=0.01MG)	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	05/20/24			
Tester	AFRE			
Recorded	05/20/24			
PPM FLUORIDE		10.0	9.8 - 10.2	9.9
		PPM		
Sample ID				
Tested	05/20/24			
Tester	AFRE			
Recorded	05/20/24			
QCSP45:MEASURING F	FLUORIDE IN	YES		YES
WATER		QCSP45		
Sample ID				
Tested	05/20/24			
Tester	AFRE			
Recorded	05/20/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID		7 <b>-</b> 2		
Tested	05/20/24			
Tester	AFRE			
Recorded	05/20/24			
INSTRUMENTS USED		INSTRUMENT		P177,M116
PREPARATION		INSTITUMENT		1 177,101110
Sample ID Tested	05/00/04			
Tested	05/20/24 AFRE			
Recorded	05/20/24			
	05/20/24	MM/DD/YYYY		11/30/2024
EXPIRATION DATE		MM/DD/YYYY		11/30/2024
Sample ID		IVIIVII DD/ 1 1 1 1		
Tested	05/20/24			
Tested	05/20/24 AFRE			
Recorded	05/20/24			
Lot # 4051722				
Made 05/20/24				



Page: 2

Date: 05/20/24 at 2:36 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**