

Page: 1

Date: 05/23/24 at 3:09 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: SPE440F			
CDTA SO	LUTION 0.02 NORMAL		
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 05/23/24			
Tester JENC			
Recorded 05/23/24			
NORMALITY TO 5 DECIMAL PLACES	0.02000	0.01990 - 0.02010	0.01997
	EQUIV/LITER		
Sample ID			
Tested 05/23/24			
Tester JENC			
Recorded 05/23/24			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
AT LANANCE AND COLOTT	APPEARANCE		TAGO
Sample ID	ALL EXITANOE		
Tested 05/23/24			
Tester JENC			
Recorded 05/23/24	INCTOLIMENT		M100 T 000 MEE DIDOC410 DID
INSTRUMENTS USED DURING	INSTRUMENT		M103,T-208,M55,BID06419,PID
PREPARATION			P-54
Sample ID			
Tested 05/23/24			
Tester JENC			
Recorded 05/23/24			
Actual Density	1.00267	1.00248 - 1.00285	1.00264
	GM/ML		
Sample ID			
Tested 05/23/24			
Tester JENC			
Recorded 05/23/24			
EXPIRATION DATE	MM/DD/YYYY		05/30/2025
	MM/DD/YYYY		
Sample ID			
Tested 05/23/24			
Tester JENC			
Recorded 05/23/24			
Lot # 4051444			
Made 05/23/24			
Made 00/20/24			



Page: 2

Date: 05/23/24 at 3:09 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager