



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: VEN001F			
Mixed ICP Standard 0.5 PPM in 5% nitric			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 05/21/24			
Tester AROS			
Recorded 05/21/24			
CALCULATED PPM CALCIUM	0.50 PPM	0.49 - 0.51	0.50
Sample ID			
Tested 05/21/24			
Tester AROS			
Recorded 05/21/24			
CALCULATED PPM COPPER (CU)	0.50 PPM	0.49 - 0.51	0.50
Sample ID			
Tested 05/21/24			
Tester AROS			
Recorded 05/21/24			
PPM IRON (FE)	0.50 PPM	0.49 - 0.51	0.50
Sample ID			
Tested 05/21/24			
Tester AROS			
Recorded 05/21/24			
CALCULATED PPM POTASSIUM	0.50 MG/L	0.49 - 0.51	0.50
Sample ID			
Tested 05/21/24			
Tester AROS			
Recorded 05/21/24			
CALCULATED PPM MAGNESIUM (MG)	0.50 PPM	0.49 - 0.51	0.50
Sample ID			
Tested 05/21/24			
Tester AROS			
Recorded 05/21/24			
CALCULATED PPM MANGANESE	0.50 PPM (UG/ML)	0.49 - 0.51	0.50
Sample ID			
Tested 05/21/24			
Tester AROS			
Recorded 05/21/24			
PPM SODIUM	0.50 PPM	0.49 - 0.51	0.50
Sample ID			
Tested 05/21/24			
Tester AROS			
Recorded 05/21/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

PPM OF LEAD	0.50	0.49 - 0.51	0.50
	PPM		
Sample ID			
Tested	05/21/24		
Tester	AROS		
Recorded	05/21/24		
CALCULATED MG/L SILICA (SI)	0.50	0.49 - 0.51	0.50
	MG/L		
Sample ID			
Tested	05/21/24		
Tester	AROS		
Recorded	05/21/24		
CALCULATED PPM ZINC (ZN)	0.50	0.49 - 0.51	0.50
	PPM		
Sample ID			
Tested	05/21/24		
Tester	AROS		
Recorded	05/21/24		
APPEARANCE AND COLOR	PASS		PASS
	APPEARANCE		
Sample ID			
Tested	05/21/24		
Tester	AROS		
Recorded	05/21/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P138,VF2-B17,GC
Sample ID			
Tested	05/21/24		
Tester	AROS		
Recorded	05/21/24		
EXPIRATION DATE	MM/DD/YYYY		05/30/2025
	MM/DD/YYYY		
Sample ID			
Tested	05/21/24		
Tester	AROS		
Recorded	05/21/24		
Lot #	4052168		
Made	05/21/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 05/21/24 at 1:50 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager