



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HON012F			
Calibration Std #3 100.0 mg/L Cl Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 05/24/24			
Tester JTRA			
Recorded 05/24/24			
PPM CHLORIDE	100.00	99.50 - 100.50	100.00
	PPM		
Sample ID			
Tested 05/24/24			
Tester JTRA			
Recorded 05/24/24			
APPEARANCE AND COLOR	CLEAR		PASS
	APPEARANCE		
Sample ID			
Tested 05/24/24			
Tester JTRA			
Recorded 05/24/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P101,VF1-Z45
Sample ID			
Tested 05/24/24			
Tester JTRA			
Recorded 05/24/24			
EXPIRATION DATE	MM/DD/YYYY		08/30/2024
	MM/DD/YYYY		
Sample ID			
Tested 05/24/24			
Tester JTRA			
Recorded 05/24/24			
Lot #	4052448		
Made	05/24/24		

Richard [Signature]

Quality Manager