Page: 1

Date: 05/21/24 at 1:27 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: UCT126	δF		
•		E STD FOR TSS 40 MG/L SOLU	TION	
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID	)			
Tested	05/21/24			
Tester	MMAT			
Recorded	05/21/24			
ACTUAL PPM		40.00	39.50 - 40.50	40.08
		PPM		
Sample ID	)			
Tested	05/21/24			
Tester	MMAT			
Recorded				
APPEARANCE AND COLOR		SLIGHTLY TURBID LIQUID		PASS
		APPEARANCE		
Sample ID	)			
Tested	05/21/24			
Tester	MMAT			
Recorded				
INSTRUMENTS USED		INSTRUMENT		M108,VF4-N9
PREPARATION				
Sample ID	)			
Tested	05/21/24			
Tester	MMAT			
Recorded				
EXPIRATION DATE	33/2 1/27	MM/DD/YY		11/30/24
LAI INATION DATE		MM/DD/YYYY		11/00/24
Sample ID	`	WIIWI/OO/1111		
Tested				
Tested	05/21/24 MMAT			
Recorded				
<b>Lot #</b> 4052				
Made 05/2	1/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St