

Page: 1

Date: 05/20/24 at 3:21 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	4704			
		0.0473 NORMAL		
Test		Target/UOM	Range	Result
PREPARED TO FORM	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	05/20/24			
Tester	SHAR			
Recorded	05/20/24			
APPEARANCE AND COLOR		CLEAR BROWN LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	05/20/24			
Tester	SHAR			
Recorded	05/20/24			
INSTRUMENTS USED DURING		INSTRUMENT		M87,T142,M55
PREPARATION				
Sample ID				
Tested	05/20/24			
Tester	SHAR			
Recorded	05/20/24			
ACTUAL NORMALITY	TO 5 PLACES	0.04730	0.04727 - 0.04733	0.04730
		EQUIV/LITER		
Sample ID				
Tested	05/20/24			
Tester	SHAR			
Recorded	05/20/24			
PREPARED TO ASTM	D 1510	YES		YES
		YES		
Sample ID				
Tested	05/20/24			
Tester	SHAR			
Recorded	05/20/24			
Actual Density		1.04326	1.03109 - 1.05543	1.04399
		GM/ML		
Sample ID				
Tested	05/20/24			
Tester	SHAR			
Recorded	05/20/24			
EXPIRATION DATE		MM/DD/YYYY		05/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	05/20/24			
Tester	SHAR			
Recorded	05/20/24			
Lot # 4051	456			
Made 05/2				
WIAUE 5572	-,			



Page: 2

Date: 05/20/24 at 3:21 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager