05/30/24

Made

Page: 1

Date: 05/30/24 at 3:27 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package		VUL393F			
	REAGENT		·	KE FRESH)	
Test			Target/UOM	Range	Result
	TO FORMULAT	ION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested 05/	30/24			
	Tester BF	RA			
	Recorded 05/	30/24			
PH @ 25 DE	EG C		9.000	8.980 - 9.020	8.994
			PH		
	Sample ID				
	•	30/24			
		RA			
	Recorded 05/	30/24			
APPEARANCE AND COLOR		C	LEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	•	30/24			
	Tester BF				
	Recorded 05/	30/24			
INSTRUMENTS USED DURING		NG	INSTRUMENT		T-52A,M108,M103,M106
PREPARAT	TON				
	Sample ID				
	•	/30/24			
		RA			
		/30/24			
EXPIRATIO			MM/DD/YYYY		08/04/2024
			MM/DD/YYYY		
	Sample ID				
	•	/30/24			
		RA			
		/30/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1.